



DOCUMENTATION SUPPLÉMENTAIRE

RAPPORT TEST NO.:

DE VÉRIFICATION ET CONTRÔLE DE QUALITÉ RÉALISÉE
PAR NOTRE SOCIÉTÉ DE GUANGZHOU (CHINE) SITUÉ À
SHENZHEN ACADEMY OF METROLOGY &
QUALITY INSPECTION.



Cet organisme de vérification est reconnu internationalement et homologué par la CNAS, organisme accrédité par le gouvernement.

**MÊME SI CE N'EST PAS OBLIGATOIRE,
NOTRE SOCIÉTÉ REVÉRIFIE QUE LA MARCHANDISE
EST RÉELLEMENT AUTHENTIQUE EN SE CONFORMANT
AUX NORMES LES PLUS EXIGEANTES D'EUROPE.**

Nous collaborons avec

Página 0 de 5



Reconocimiento
a la responsabilidad
ambiental



Universidad
Politécnica
de Cartagena



UCAM
UNIVERSIDAD
CATÓLICA DE MURCIA

UIMPCartagena
Universidad Internacional Menéndez Pelayo



Associations et entités auxquelles nous appartenons





TEST REPORT

Lot No. / Batch Code:
 Sample Description:
 Quantity Submitted: 40PCS
 Manufacturer / Supplier:
 Manufactured Date: /
 Sample Receiving Condition: In unopened plastic bag under ambient condition
 Country of Origin:
 Sample Receiving Date: Sep. 10, 2020
 Testing Period: Sep. 10, 2020 –Sep.16, 2020

Test Requested : Please refer to the result summary.

Test Method & Results : Please refer to next page(s).

Result Summary :

Test Requested	Conclusion
EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods	PASS

Signed for and on behalf of Shenzhen Academy of Metrology & Quality Inspection:

Approver

Checker

Technical Director



TEST REPORT

EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

<u>Clause</u>	<u>Test Items/requirement</u>	<u>Test Result Summary</u>
5	Requirements	
5.2	Performance requirements	
5.2.2	Bacterial filtration efficiency (BFE)	> 98%
5.2.6	Summary of performance requirements	See Table 1

Table 1 Performance requirements for medical face masks

Characteristics	Type I ^a	Type II	Type IIR
Bacterial filtration efficiency (BFE), %	≥ 95	≥ 98	≥ 98
Differential pressure, Pa/cm ²	< 40	< 40	< 60
Splash resistance (kPa) [#]	Not Required	Not Required	≥16.0
Microbial cleanliness (cfu/g)	≤ 30	≤ 30	≤ 30

^a Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

[#] - An acceptable quality limit of 4,0 % is met for a single sampling plan when 29 or more of the 32 tested specimens show “pass” results.



TEST REPORT

Result 1 Bacterial filtration efficiency (BFE) (EN14683:2019+AC:2019 Appendix B)

Blue mask:

Test Side: White Colour (Inside)
Pre-Conditioning: 5 hours at 21 °C±5°C and 85%±5% R.H.
Dimensions of test specimen: 14cm x 12cm
BFE Test Area: 95 cm²
BFE Flow Rate: 28.3 l/min
Test bacteria: Staphylococcus aureus ATCC 6538
Positive Control Average: 2065CFU
Negative Monitor Count: 0 CFU

Test Specimen	Percent BFE (%)
1	99.71
2	99.76
3	99.61
4	99.23
5	99.76



TEST REPORT

Black mask:
Test Side: Black Colour (Inside)
Pre-Conditioning: 5 hours at 21 °C±5°C and 85%±5% R.H.
Dimensions of test specimen: 14cm x 12cm
BFE Test Area: 95 cm²
BFE Flow Rate: 28.3 l/min
Test bacteria: Staphylococcus aureus ATCC 6538
Positive Control Average: 2065CFU
Negative Monitor Count: 0 CFU

Test Specimen	Percent BFE (%)
1	99.85
2	99.90
3	99.85
4	99.81
5	99.90

TEST REPORT

Note:

- 1. Results reported on the submitted sample on an as received basis.
- 2. The analysis was performed by a SMQ assessed competent subcontractor laboratory.

Photo Appendix



***** End of Report *****